



Please refer to the **TENDER FOR SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF VIAL PROCESSING UNIT AT KABETE PRODUCTION UNIT, Ref No. KVVPI-IT-01- 2019-2020** that was advertised on 12th June 2020.

We hereby issue this 2nd addendum giving clarification/amending the tender document as follows:

Clarification/Amendment		KEVEVEPI's clarification/amendment to the tender.	
1.	Please consider an extension of the tender closing date to be on 17.07. 2020, and The reason is the movement challenges as a result of the Covid-19 restrictions worldwide. 03rd July, 2020 at 12.00pm EAT)	1.	New opening date amended to 17th July, 2020 at 12.00pm E.A.T
2.	Sec. 3.18.1 Liquidated damages Please review the maximum deduction from 10% to be 5%. The tender should be given 2 months' notice prior to termination	2.	<i>3.18.1 If the tenderer fails to deliver and/or install any or all of the items within the period(s) specified in the contract, the procuring entity shall, without prejudice to its other remedies under the contract, deduct from the contract prices liquidated damages sum equivalent to 0.5% of the delivered price of the delayed items up to a maximum deduction of 5% of the delayed goods. Tenderers will be issued with three-month notice period, after this the tenderer may consider termination of the contract.</i>

Clarification/Amendment		KEVEVEPI's clarification/amendment to the tender.	
3.	<p>Section 3.8.4 Inspection and Tests</p> <p>Kindly note that we ask you to delete sec 3.8.4 The reason is that the machinery be tested twice in FAT test in our Factory and in SAT test also and instillation in Kenya , so as per contract will not be need for that as that test should be already done in the Fat test and signed with acceptance in FAT test by Kevevepi engineers and our engineers (signed and confirmed after the machine pass all tests in FAT test in the presence of all Kevevepi Team)</p>	3.	<p><i>Section 3.8.4 is deleted</i></p> <ul style="list-style-type: none"> <i>However, inspection and tests will be undertaken during the SAT and FAT by Kevevapi staff. Tenderers should note the requirements stated in the technical specifications will be applicable: WARRANTY: 2 YEARS against manufacturer's defects</i>

CHIEF EXECUTIVE OFFICER